





Acq. Method : D:\METHODS\1-POS-3MIN.M  
 Analysis Method : D:\METHODS\1-POS-3MIN.M  
 Sample Info : Easy-Access Method: '1-POS-3MIN'  
 Method Info : Mobile Phase: A: water(0.01%TFA) B: ACN(0.01%TFA)  
 Gradient: 5% increase to 95%B within 1.3 min, 95%B for 1.7 min.  
 Flow Rate: 2.0ml/min  
 Column:Sunfire C18,4.6\*50mm,3.5um  
 Column Temperature: 45 C  
 Detection:UV(214,4nm) and MS(ESI,Pos mode,110 to 1000 amu)



Integration Results for ADC1 A, ELSD

RetTim	Width	Area	Height	Area%
1.22	0.06	162.70	39.05	100.00


Integration Results for DAD1 A, Sig=214,4 Ref=off

RetTim	Width	Area	Height	Area%
1.21	0.07	5772.02	1114.13	100.00


Integration Results for DAD1 B, Sig=254,4 Ref=off

RetTim	Width	Area	Height	Area%
1.21	0.07	4278.43	829.46	100.00


Ret. Time: 0.30 <<<< POSITIVE SPECTRA >>>>



Ret. Time: 1.11



Ret. Time: 1.22



Ret. Time: 1.85

